



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

1 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No               | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                     | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|--------------------|---|---|--|---|--|
| Permanent Facility |   |   |  |   |  |
| 1                  | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM ,By Direct Method                        | 1 mA to 10 mA   | 0.16%  |
| 2                  | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM ,By Direct Method                        | 10 µA to 1 mA   | 0.85 % to 0.16 %                                 |
| 3                  | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM ,By Direct Method                        | 10 mA to 10 A   | 0.16 % to 0.25 %                                 |
| 4                  | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current (1 Phase) (At 50 Hz)   | Using HIOKI Digital Power Meter/Harmonic Analyzer,By Direct Method | 10 mA to 65 A   | 0.06%  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

2 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---|---|---|---|--|
| 5    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current (3 Phase) (At 50Hz)  | Using HIOKI Digital Power Meter/Harmonic Analyzer ,By Direct Method | 10 mA to 50 A   | 0.13 % to 0.034 %                                |
| 6    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)  | AC Current (1 Phase & 3 Phase) @ 50 Hz  | Using CALMET C300 Power/Energy Calibrator by Direct Method          | 20 A to 120 A   | 0.3%   |
| 7    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)  | AC Current (1 Phase) @ 50 Hz  | Using Fluke 9100 with Current coil ,By Direct Method                | 20 A to 1000 A  | 0.76%  |
| 8    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)  | AC Current (1 Phase) @ 50Hz to 5kHz   | Using Fluke 9100 MFC ,By Direct Method                              | 10 $\mu$ A to 300 $\mu$ A   | 4.1 % to 0.25 %                                  |
| 9    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)  | AC Current (1 Phase) @ 50Hz to 5kHz   | Using Fluke 9100 MFC ,By Direct Method                              | 300 mA to 20 A  | 0.25 % to 0.3 %                                  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

3 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                                       | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--|---|---|---|--|
| 10   | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source) | AC Current(1 Phase) @ 50Hz to 5kHz  | Using Fluke 9100 MFC ,By Direct Method                              | 300 $\mu$ A to 300 mA   | 0.25%  |
| 11   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)               | DC CURRENT  | Using Fluke 8846A, 6½ DMM ,By Direct Method                         | 0.10 mA to 10 mA  | 0.085 % to 0.065 %                               |
| 12   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)               | DC CURRENT  | Using Fluke 8846A, 6½ DMM ,By Direct Method                         | 1 $\mu$ A to 100 $\mu$ A  | 3.06 % to 0.085 %                                |
| 13   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)               | DC CURRENT  | Using HIOKI Digital Power Meter/Harmonic Analyzer ,By Direct Method | 10 A to 65 A  | 0.13 % to 0.12 %                                 |
| 14   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)               | DC CURRENT  | Using Fluke 8846A, 6½ DMM ,By Direct Method                         | 10 mA to 10 A   | 0.065 % to 0.8 %                                 |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

4 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                              | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---|---|---|---|--|
| 15   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC CURRENT  | Using Fluke 9100 MFC, By Direct Method                                      | 10 $\mu$ A to 300 mA  | 0.59 % to 0.035 %                                |
| 16   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC CURRENT  | Fluke 9100 with current coils ,By Direct Method                             | 20 A to 1000 A  | 0.8 % to 0.8 %                                   |
| 17   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC CURRENT  | Using Fluke 9100 MFC, By Direct Method                                      | 300 mA to 20 A  | 0.035%   |
| 18   | ELECTRO-TECHNICAL-OTHERS (Measure)        | AC High Voltage (At 50Hz)   | Using H.V. Probe & DMM Fluke ,By Direct Method                              | 1 kV to 28 kV   | 2.33%  |
| 19   | ELECTRO-TECHNICAL-OTHERS (Measure)        | AC Power (1 Phase) (At 50Hz) At UPF (10V to 600 V 0.1A to 65 A)   | Using Yokogawa/HIOKI Digital power Meter/Harmonic Analyzer,By Direct Method | 0.1 W to 39 kW  | 0.9%   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

5 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|---|---|--|
| 20   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Power (3 Phase) (At 50Hz ,At UPF) (10 V to 600 V, 0.01A to 50 A)   | Using HIOKI Digital Power Meter/Harmonic Analyzer,By Direct Method  | 0.1 W to 30 kW  | 0.13 % to 0.034 %                                |
| 21   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Voltage (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM , By Direct Method                        | 1 mV to 100 mV  | 4.7 % to 0.12 %                                  |
| 22   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Voltage (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM , By Direct Method                        | 1 V to 1000 V   | 0.1%   |
| 23   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Voltage (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM , By Direct Method                        | 100 mV to 1 V   | 0.12 % to 0.1 %                                  |
| 24   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Voltage (1 Phase) (At 100 kHz)   | Using Fluke 8846A, 6½ DMM , By Direct Method                        | 10 mV to 100 V  | 6 % to 0.8 %                                     |
| 25   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Voltage (3 Phase) (At 50Hz)  | Using HIOKI Digital Power Meter/Harmonic Analyzer ,By Direct Method | 0.1 V to 1000 V   | 0.7 % to 0.06 %                                  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

6 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                     | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|--|---|--|
| 26   | ELECTRO-TECHNICAL-OTHERS (Measure) | Capacitance (At 1 kHz)  | Using LCR Meter ,By Direct Method                                  | 1 $\mu$ F to 1011 $\mu$ F   | 0.91 % to 0.31 %                                 |
| 27   | ELECTRO-TECHNICAL-OTHERS (Measure) | Capacitance (At 1 kHz)  | Using LCR Meter ,By Direct Method                                  | 1 nF to 1000 nF   | 0.91 % to 0.06 %                                 |
| 28   | ELECTRO-TECHNICAL-OTHERS (Measure) | Capacitance (At 1 kHz)  | Using LCR Meter ,By Direct Method                                  | 1 pF to 1000 pF   | 0.96 % to 0.06 %                                 |
| 29   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC High Voltage   | Using H.V. Probe & DMM Fluke ,By Direct Method                     | 1 kV to 40 kV   | 2.38%  |
| 30   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC Power (10 V to 600 V, 0.01A to 65 A)   | Using HIOKI Digital Power Meter/Harmonic Analyzer,By Direct Method | 0.1 W to 39 kW  | 0.2%   |
| 31   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC Voltage  | Using Fluke 8846A, 6½ DMM ,By Direct Method                        | 0.1 mV to 1 mV  | 4.0 % to 0.4 %                                   |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

7 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|--|---|--|
| 32   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC Voltage  | Using Fluke 8846A, 6½ DMM ,By Direct Method                                  | 1 mV to 100 mV  | 0.4 % to 0.0085 %                                |
| 33   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC Voltage  | Using Fluke 8846A, 6½ DMM , By Direct Method                                 | 1 V to 1000 V   | 0.0045 % to 0.006 %                              |
| 34   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC Voltage  | Using Fluke 8846A, 6½ DMM , By Direct Method                                 | 100 mV to 1 mV  | 0.0085 % to 0.0045 %                             |
| 35   | ELECTRO-TECHNICAL-OTHERS (Measure) | Harmonics (1 Phase,3 Phase) (200V, 10A, 50Hz)   | Using Yokogawa/HIOKI Digital power Meter/Harmonic Analyzer ,By Direct Method | 3 <sup>rd</sup> to 49 <sup>th</sup> Order   | 0.51%  |
| 36   | ELECTRO-TECHNICAL-OTHERS (Measure) | Harmonics (1 Phase,3 Phase) (200V-240V, 0.5A to 10A, 50Hz)  | Using Yokogawa/HIOKI Digital power Meter/Harmonic Analyzer ,By Direct Method | 3 <sup>rd</sup> to 39 <sup>th</sup> Order   | 0.51%  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

8 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|--|---|--|
| 37   | ELECTRO-TECHNICAL-OTHERS (Measure) | Impulse (Duration - Half pulse width)   | Using Oscilloscope with HV Probe by Direct/Comparison method as per IEC 60060-2, IEC 61180, IS 2071 Part 1 | 10 $\mu$ S to 1000 $\mu$ S  | 1.6 % to 2.5 %                                   |
| 38   | ELECTRO-TECHNICAL-OTHERS (Measure) | Impulse (Front Time)  | Using Oscilloscope with HV Probe by Direct/Comparison method as per IEC 60060-2, IEC 61180, IS 2071 Part 1 | 0.8 $\mu$ S to 1000 $\mu$ S   | 1.6 % to 2.5 %                                   |
| 39   | ELECTRO-TECHNICAL-OTHERS (Measure) | Impulse (Voltage Magnitude)   | Using Oscilloscope with HV Probe by Direct/Comparison method as per IEC 60060-2, IEC 61180, IS 2071 Part 1 | 1 kVp to 17 kVp   | 3.69%  |
| 40   | ELECTRO-TECHNICAL-OTHERS (Measure) | Inductance (At 1kHz)  | Using LCR Meter ,By Direct Method  | 1 H to 10 H   | 0.061 % to 0.06 %                                |
| 41   | ELECTRO-TECHNICAL-OTHERS (Measure) | Inductance (At 1kHz)  | Using LCR Meter ,By Direct Method  | 1 mH to 1003 mH   | 0.06 % to 0.28 %                                 |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

9 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|--|---|--|
| 42   | ELECTRO-TECHNICAL-OTHERS (Measure) | Inductance (At 1kHz)  | Using LCR Meter ,By Direct Method  | 99 $\mu$ H to 999.9 $\mu$ H   | 0.28%  |
| 43   | ELECTRO-TECHNICAL-OTHERS (Measure) | Power Factor  | Using Yokogawa/HIOKI Digital power Meter/Harmonic Analyzer, By Direct Method | 0.1 1PF lag - UPF to 0.1 PF lead  | 0.007PF  |
| 44   | ELECTRO-TECHNICAL-OTHERS (Measure) | Resistance  | Using LCR Meter (At 1kHz), By Direct Method                                  | 0.01 Ohm to 100 Ohm   | 0.6 % to 0.07 %                                  |
| 45   | ELECTRO-TECHNICAL-OTHERS (Measure) | Resistance  | Using Fluke 8846A, 6½ DMM,By Direct Method                                   | 0.1 Ohm to 100 kOhm   | 3.6 % to 0.013 %                                 |
| 46   | ELECTRO-TECHNICAL-OTHERS (Measure) | Resistance  | Using Fluke 8846A, 6½ DMM,By Direct Method                                   | 10 MOhm to 1 GOhm   | 0.05 % to 2.5 %                                  |
| 47   | ELECTRO-TECHNICAL-OTHERS (Measure) | Resistance  | Using Fluke 8846A, 6½ DMM, By Direct Method                                  | 100 kOhm to 10 MOhm   | 0.013 % to 0.05 %                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

10 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|--|---|--|
| 48   | ELECTRO-TECHNICAL-OTHERS (Measure) | Surge   | Using Oscilloscope with HV Probe by Direct/Comparison method as per IEC 60060-2, IEC 61180, IS 2071 Part 1 | . up to 10 kV   | 1.6 % to 3.69 %                                  |
| 49   | ELECTRO-TECHNICAL-OTHERS (Source)  | AC Energy (1Phase & 3Phase) (Active/Reactive) UPF to 0.25PF Lead/Lag @ 50Hz (63.5 V to 240 V, 0.05 A to 120 A)              | Using CALMET C300 Power/Energy calibrator ,By comparison method  | 0.79 Wh to 28.8 kWh   | 0.68 % to 0.06 %                                 |
| 50   | ELECTRO-TECHNICAL-OTHERS (Source)  | AC Power (1 Phase) (P.F. 0.8 Lag to 0.8 Lead) (At 50Hz)   | Using Fluke 9100 MFC,By Direct Method  | 2.4 W to 8.96 kW  | 0.08 % to 0.65 %                                 |
| 51   | ELECTRO-TECHNICAL-OTHERS (Source)  | AC Power (1 Phase) P.F. 0.5 Lag to 0.5 Lead (At 50Hz )  | Using Fluke 9100 MFC ,By Direct Method   | 1.5 W to 5.6 kW   | 0.08 % to 0.65 %                                 |
| 52   | ELECTRO-TECHNICAL-OTHERS (Source)  | AC Power (1 Phase) UPF @ 50Hz (30 V to 560 V, 0.1A to 20 A)   | Using Fluke 9100 MFC ,By Direct Method   | 3 W to 11.2 kW  | 0.06 % to 0.6 %                                  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

11 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|---|--|---|--|
| 53   | ELECTRO-TECHNICAL-OTHERS (Source) | AC Power (3 Phase) (UPF) @ 50Hz (10 V to 500 V, 0.1A to 100 A)  | Using CALMET C300 Power/Energy calibrator,By Direct Method   | 1 W to 50.0 kW  | 1.5%   |
| 54   | ELECTRO-TECHNICAL-OTHERS (Source) | AC Voltage (1 Phase) (At 10Hz to 100 KHz )  | Using Fluke 9100 MFC , By Direct Method                      | 10 mV to 32 mV  | 4.56 % to 0.55 %                                 |
| 55   | ELECTRO-TECHNICAL-OTHERS (Source) | AC Voltage (1 Phase) (10Hz to 100 kHz)  | Using Fluke 9100 MFC , By Direct Method                      | 32 mV to 320 mV   | 0.55 % to 0.25 %                                 |
| 56   | ELECTRO-TECHNICAL-OTHERS (Source) | AC Voltage (1 Phase) (10Hz to 100kHz)   | Using Fluke 9100 MFC , By Direct Method                      | 320 mV to 1000 V  | 0.25 % to 0.2 %                                  |
| 57   | ELECTRO-TECHNICAL-OTHERS (Source) | AC Voltage (3 Phase) (At 50 Hz)   | Using CALMET C300 Power/ Energy Calibrator ,By Direct Method | 0.5 V to 560 V  | 0.043 % to 0.049 %                               |
| 58   | ELECTRO-TECHNICAL-OTHERS (Source) | Capacitance (At 1 kHz)  | Using Fluke 9100 MFC , By Direct Method                      | 1 nF to 100 nF  | 2.04 % to 0.55 %                                 |
| 59   | ELECTRO-TECHNICAL-OTHERS (Source) | Capacitance (At 1 kHz)  | Using Fluke 9100 MFC , By Direct Method                      | 100 nF to 1 mF  | 0.55 % to 1.85 %                                 |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

12 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|---|--|---|--|
| 60   | ELECTRO-TECHNICAL-OTHERS (Source) | DC Power (1 Phase) (10 V to 600 V, 0.1A to 20 A)  | Using Fluke 9100 MFC,By Direct Method                            | 1 W to 12 kW  | 6.3 % to 0.2 %                                   |
| 61   | ELECTRO-TECHNICAL-OTHERS (Source) | DC Voltage  | Using Fluke 9100 MFC , By Direct Method                          | 1 mV to 300 mV  | 0.7 % to 0.055 %                                 |
| 62   | ELECTRO-TECHNICAL-OTHERS (Source) | DC Voltage  | Using Fluke 9100 MFC , By Direct Method                          | 100 V to 1000 V   | 0.015 % to 0.01 %                                |
| 63   | ELECTRO-TECHNICAL-OTHERS (Source) | DC Voltage  | Using Fluke 9100 MFC , By Direct Method                          | 300 mV to 100 V   | 0.055 % to 0.015 %                               |
| 64   | ELECTRO-TECHNICAL-OTHERS (Source) | High Resistance# (I.R Option) (at 250V to 1kV)  | Using Fluke 9100 MFC , By Direct Method                          | 1 MOhm to 1.8 GOhm  | 0.2 % to 1.1 %                                   |
| 65   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 0.1 mOhm to 0.1 mOhm  | 0.96%  |
| 66   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 1 mOhm to 1 mOhm  | 0.79%  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

13 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|---|--|---|--|
| 67   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 1 Ohm to 1 Ohm  | 0.31%  |
| 68   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 10 mOhm to 10 mOhm  | 0.31%  |
| 69   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 100 mOhm to 100 mOhm  | 0.31%  |
| 70   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 2 Ohm to 2 Ohm  | 0.31%  |
| 71   | ELECTRO-TECHNICAL-OTHERS (Source) | Power Factor @ 50Hz   | Using Fluke 9100 MFC by Direct Method                            | 0.1 PF lag -UPF to 0.1 PF lead  | 0.002 % to 0.001 %                               |
| 72   | ELECTRO-TECHNICAL-OTHERS (Source) | Resistance (2W)   | Using Fluke 9100 MFC ,By Direct Method                           | 1 MOhm to 10 MOhm   | 0.076 % to 0.20 %                                |
| 73   | ELECTRO-TECHNICAL-OTHERS (Source) | Resistance (2W)   | Using Fluke 9100 MFC ,By Direct Method                           | 10 Ohm to 100 kOhm  | 0.20 % to 0.035 %                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

14 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--|---|---|---|--|
| 74   | ELECTRO-TECHNICAL-OTHERS (Source)                  | Resistance (2W)   | Using Fluke 9100 MFC ,By Direct Method              | 10 MOhm to 300 MOhm   | 0.02 % to 0.59 %                                 |
| 75   | ELECTRO-TECHNICAL-OTHERS (Source)                  | Resistance (2W)   | Using Fluke 9100 MFC ,By Direct Method              | 100 kOhm to 1 MOhm  | 0.032 % to 0.076 %                               |
| 76   | ELECTRO-TECHNICAL-OTHERS (Source)                  | RESISTANCE (2W)   | Using Fluke 9100 MFC ,By Direct Method              | 3 Ohm to 10 Ohm   | 0.46 % to 0.20 %                                 |
| 77   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | Temperature Simulation (By simulation method) (J Type Thermocouple)   | Using Cropico digital thermometer, By Direct Method | -200 °C to 1200 °C  | 0.55°C   |
| 78   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | Temperature Simulation (By simulation method) (K Type Thermocouple))  | Using Cropico digital thermometer, By Direct Method | -200 °C to 1300 °C  | 0.52°C   |
| 79   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | Temperature Simulation (By simulation method) (RTD Type)  | Using Cropico digital thermometer, By Direct Method | -200 °C to 600 °C   | 0.2°C  |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

15 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---|---|--|---|--|
| 80   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | Harmonics (1Phase & 3Phase) (10V to 300V, 0.5A to 100A)   | Using CALMET C300 Power/Energy Calibrator , By Direct Method | 2 <sup>nd</sup> to 31 <sup>st</sup>   | 0.6%   |
| 81   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | Temperature Indicator, Controller, Recorders (By simulation method) (J Type Thermocouple)                                   | Using Fluke 9100 MFC, By Direct Method                       | -200 °C to 1200 °C  | 0.3°C  |
| 82   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | Temperature Indicator, Controller, Recorders (By simulation method) (RTD TYPE)  | Using Fluke 9100 MFC, By Direct Method                       | -200 °C to 800 °C   | 0.2°C  |
| 83   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | Temperature Indicator, Controller, Recorders (By simulation method) (K Type Thermocouple)                                   | Using Fluke 9100 MFC, By Direct Method                       | -200 °C to 1300 °C  | 0.3°C  |
| 84   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | FREQUENCY (1 Phase)   | Using Fluke 8846A, 6½ DMM ,By Direct Method                  | 10 Hz to 1 MHz  | 0.17 % to 0.011 %                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

16 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                           | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--|---|---|---|--|
| 85   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure) | FREQUENCY (3 Phase) (At 240V)   | Using HIOKI Digital Power Meter/Harmonic Analyzer,By Direct Method  | 50 Hz to 50 kHz   | 0.02%  |
| 86   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure) | TIME INTERVAL   | Using Time Interval Meter & Digital Stop Watch,By Comparison Method | 0.1 s to 999.99 s   | 0.002 Sec to 1.13 Sec                            |
| 87   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure) | TIME INTERVAL   | Using Time Interval Meter & Digital Stop Watch,By Comparison Method | 1000 s to 24 hrs  | 0.002 sec to 1.61 sec                            |
| 88   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)  | FREQUENCY   | Using Fluke 9100 MFC , By Direct Method                             | 1 Hz to 100 Hz  | 0.6 % to 0.006 %                                 |
| 89   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)  | FREQUENCY   | Using Fluke 9100 MFC , By Direct Method                             | 100 Hz to 1 MHz   | 0.006 % to 0.003 %                               |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

17 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---|---|--|---|--|
| 90   | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Calipers (Vernier / Digital /Dial) L.C.: 0.01 mm  | Using Gauge Block / Caliper Checker by Comparison method                                 | 0 to 300 mm   | 13µm   |
| 91   | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | External Micrometer L.C.: 0.001 mm  | Using Gauge Block Set by Comparison method   | 0 to 100 mm   | 2.7µm  |
| 92   | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Feeler Gauge  | Using Digital Micrometer by Comparison method  | 0.01 mm to 2 mm   | 2.7µm  |
| 93   | MECHANICAL-PRESSURE INDICATING DEVICES                        | (Pneumatic-Gauge Pressure) Digital & Dial Pressure Gauge  | Using Digital Pressure Gauge & Comparator by Comparison Method as per DKD- R - 6 -1      | 0 to 10 bar   | 0.20% of rdg                                     |
| 94   | MECHANICAL-PRESSURE INDICATING DEVICES                        | (Pneumatic-Gauge Pressure) Digital & Dial Pressure Gauge  | Using Digital Vacuum Gauge & Vacuum Comparator by Comparison Method as per DKD- R - 6 -1 | -90 bar to 0 bar  | 0.01% of rdg                                     |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

18 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                     | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--|---|---|---|--|
| 95   | MECHANICAL-PRESSURE INDICATING DEVICES | Hydraulic Pressure ( Analog / Digital Gauges)   | Using Dead weight Tester by Comparison Method as per DKD- R - 6 -1  | 0.1 bar to 700 bar  | 0.076% of rdg                                    |
| 96   | MECHANICAL-PRESSURE INDICATING DEVICES | Hydraulic Pressure # ( Analog / Digital Gauges)   | Using Digital Pressure Gauge with Hydraulic Comparator by Comparison Method as per DKD- R - 6 -1                            | 0 to 700 bar  | 0.1% of rdg                                      |
| 97   | MECHANICAL-VOLUME                      | Glassware like pipettes, burettes, measuring cylinder, volumetric flask etc @ 27°C  | Using Weighing balance of 10kg capacity and 0.1g readability and distilled water ,Gravimetric method based on IS/ISO 4787   | >2000 ml to 5000 ml   | 1.33ml   |
| 98   | MECHANICAL-VOLUME                      | Glassware like pipettes, burettes, measuring cylinder, volumetric flask etc @ 27°C  | Using Weighing balance of 200g capacity and 0.1mg readability and distilled water , Gravimetric method based on IS/ISO 4787 | >50 ml to 100 ml  | 0.5ml  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

19 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 99   | MECHANICAL-VOLUME  | Glassware like pipettes, burettes, measuring cylinder, volumetric flask etc @ 27°C  | Using Weighing balance of 200g capacity and 0.1mg readability and distilled water , Gravimetric method based on IS/ISO 4787          | 1 ml to 50 ml   | 0.1ml  |
| 100  | MECHANICAL-VOLUME  | Micro-pipette @ 27°C  | Using Weighing balance of 200g capacity and 0.1mg readability and distilled water , Gravimetric method based on ISO 8655 part6       | 100 µl to 200 µl  | 2.1µl  |
| 101  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 1 g   | 0.11mg   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

|                               |   |                        |          |
|-------------------------------|---|------------------------|----------|
| <b>Laboratory Name :</b>      | HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA |                        |          |
| <b>Accreditation Standard</b> | ISO/IEC 17025:2017  |                        |          |
| <b>Certificate Number</b>     | CC-2165   | <b>Page No</b>         | 20 of 64 |
| <b>Validity</b>               | 03/12/2018 to 02/12/2020*   | <b>Last Amended on</b> | -        |

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 102  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 1 mg  | 0.11mg   |
| 103  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 10 g  | 0.11mg   |
| 104  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 10 mg   | 0.11mg   |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

21 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 105  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 100 g   | 0.12mg   |
| 106  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 100 mg  | 0.11mg   |
| 107  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 2 g   | 0.11mg   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

22 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 108  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 2 mg  | 0.11mg   |
| 109  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 20 g  | 0.12mg   |
| 110  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 20 mg   | 0.11mg   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

23 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 111  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 200 g   | 0.16mg   |
| 112  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 200 mg  | 0.11mg   |
| 113  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 5 g   | 0.11mg   |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

24 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 114  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 5 mg  | 0.11mg   |
| 115  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 50 g  | 0.12mg   |
| 116  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 50 mg   | 0.11mg   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

25 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|---|---|--|
| 117  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111  | 500 mg  | 0.11mg   |
| 118  | OPTICAL-OPTICAL    | Chromaticity Coordinates of LED Lamp  | Using C-Type mirror Goniometer with Spectroradiometer and integrating sphere with standard spectroradiometer/ Spectrophotometer and Standard light source by Obsolete & Substitution method | x =0.005  | 1.0%   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

26 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|---|---|--|
| 119  | OPTICAL-<br>OPTICAL | Chromaticity Coordinates of LED Lamp  | Using C-Type mirror Goniometer with Spectroradiometer and integrating sphere with standard spectroradiometer/ Spectrophotometer and Standard light source by Obsolete & Substitution method | y =0.005  | 1.0%   |
| 120  | OPTICAL-<br>OPTICAL | Color Coordinates   | Using THL lamps at 50W & integrating sphere with standard spectrophotometer by ab-initio Method & Substitution method   | x =0.005  | 1.0%   |
| 121  | OPTICAL-<br>OPTICAL | Color Coordinates   | Using THL lamps at 50W & integrating sphere with standard spectrophotometer by ab-initio Method & Substitution method   | y =0.005  | 1.0%   |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

27 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|---|---|--|
| 122  | OPTICAL-<br>OPTICAL | Compact Fluorescent Lamps (CFL) (5 W to 26 W)   | Using standard lamps & integrating sphere with standard spectrophotometer by ab-initio Method & Substitution method   | 200 lm to 1600 lm   | 3.0%   |
| 123  | OPTICAL-<br>OPTICAL | Correlated Color Temperature  | Using THL lamps at 50W & integrating sphere with standard spectrophotometer by ab-initio Method & Substitution method   | 1500 K to 25000 K   | 1.5%   |
| 124  | OPTICAL-<br>OPTICAL | Correlated Color Temperature of LED Lamp  | Using C-Type mirror Goniometer with Spectroradiometer and integrating sphere with standard spectroradiometer/ Spectrophotometer and Standard light source by Obsolete & Substitution method | 1500 K to 25000 K   | 1.2%   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

|                               |   |                        |          |
|-------------------------------|---|------------------------|----------|
| <b>Laboratory Name :</b>      | HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA |                        |          |
| <b>Accreditation Standard</b> | ISO/IEC 17025:2017  |                        |          |
| <b>Certificate Number</b>     | CC-2165   | <b>Page No</b>         | 28 of 64 |
| <b>Validity</b>               | 03/12/2018 to 02/12/2020*   | <b>Last Amended on</b> | -        |

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|--|---|--|
| 125  | OPTICAL-<br>OPTICAL | HPMVL (80W to 400W)   | Using standard lamps & integrating sphere with standard spectrophoto colorimeter by ab-initio Method & Substitution method | 2000 lm to 30000 lm   | 3.2%   |
| 126  | OPTICAL-<br>OPTICAL | HPSVL (70W to 400W)   | Using standard lamps & integrating sphere with standard spectrophoto colorimeter by ab-initio Method & Substitution method | 4800 lm to 60000 lm   | 3.2%   |
| 127  | OPTICAL-<br>OPTICAL | Illuminance   | Using Illuminance meter with Intensity Standard Lamp WI 41G Lamp by Direct Method  | 1 lux to 25000 lux  | 2.1%   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

29 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|---|---|--|
| 128  | OPTICAL-<br>OPTICAL | Luminous Flux Of LED Lamp (0.5 W to 23 W)   | Using C-Type mirror Goniometer with Spectroradiometer and integrating sphere with standard spectroradiometer/ Spectrophotometer and Standard light source by Obsolete & Substitution method | 20 lm to 3000 lm  | 2.2%   |
| 129  | OPTICAL-<br>OPTICAL | Luminous Intensity  | Using Illuminance meter with Intensity Standard Lamp WI 41G Lamp by Direct Method   | 1 cd to 10000 cd  | 2.1%   |
| 130  | OPTICAL-<br>OPTICAL | MHL (70W to 400W)   | Using standard lamps & integrating sphere with standard spectrophotometer by ab-initio Method & Substitution method   | 4800 lm to 60000 lm   | 3.5%   |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

30 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                   | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------------------------|---|---|---|--|
| 131  | OPTICAL-<br>OPTICAL                  | TFL (6W to 40W)   | Using standard lamps & integrating sphere with standard spectrophotometer   | 150 lm to 4000 lm   | 3.0%   |
| 132  | OPTICAL-<br>OPTICAL                  | Tungsten Filament Lamps (15 W to 200 W)   | Using standard lamps & integrating sphere with standard spectrophotometer by ab-initio Method & Substitution method | 115 lm to 3400 lm   | 3.0%   |
| 133  | THERMAL-<br>SPECIFIC HEAT & HUMIDITY | Calibration of Furnace, Freezer, Oven ,Chamber  | Using 16-Channel Temperature Scanner with minimum 9 RTD Sensors by Multi Position Calibration                       | 140 °C to 250 °C  | 1.7°C  |
| 134  | THERMAL-<br>SPECIFIC HEAT & HUMIDITY | Calibration of Furnace, Freezer, Oven ,Chamber  | Using 16-Channel Temperature Scanner with minimum 9 RTD Sensors by Multi Position Calibration                       | -40 °C to 140 °C  | 1.7°C  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

31 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|--|---|--|
| 135  | THERMAL-TEMPERATURE | Calibration of Contact type temp. sensor (RTDs, Thermocouples, Temp. indicators with Sensors)                               | Using Standard RTD, S- Type T/C Sensor with Standard Digital Thermometer and Low Temperature Bath & Dry Block Calibrators , By Comparison Method | 100 °C to 200 °C  | 0.14°C   |
| 136  | THERMAL-TEMPERATURE | Calibration of Contact type temp. sensor (RTDs, Thermocouples, Temp. indicators with Sensors)                               | Using Standard RTD, S- Type T/C Sensor with Standard Digital Thermometer and Low Temperature Bath & Dry Block Calibrators , By Comparison Method | 200 °C to 600 °C  | 0.32°C   |
| 137  | THERMAL-TEMPERATURE | Calibration of Contact type temp. sensor (RTDs, Thermocouples, Temp. indicators with Sensors)                               | Using Standard RTD, S- Type T/C Sensor with Standard Digital Thermometer and Low Temperature Bath & Dry Block Calibrators , By Comparison Method | -40 °C to 100 °C  | 0.11°C   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

|                               |   |                        |          |
|-------------------------------|---|------------------------|----------|
| <b>Laboratory Name :</b>      | HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA |                        |          |
| <b>Accreditation Standard</b> | ISO/IEC 17025:2017  |                        |          |
| <b>Certificate Number</b>     | CC-2165   | <b>Page No</b>         | 32 of 64 |
| <b>Validity</b>               | 03/12/2018 to 02/12/2020*   | <b>Last Amended on</b> | -        |

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|--|---|--|
| 138  | THERMAL-TEMPERATURE | Calibration of Contact type temp. sensor (Thermocouples, Temp. indicators with Sensors)                                     | Using Standard RTD, S- Type T/C Sensor with Standard Digital Thermometer and Low Temperature Bath & Dry Block Calibrators , By Comparison Method | 600 °C to 1000 °C   | 1.55°C   |
| 139  | THERMAL-TEMPERATURE | Glass Thermometer, Dial Gauge   | Using Standard RTD Sensor with Standard Digital Thermometer and Low Temperature Bath , By Comparison Method                                      | -40 °C to 200 °C  | 0.65°C   |
| 140  | THERMAL-TEMPERATURE | Temperature Indicator with sensor of Freezer, Chamber, Bath, Oven, & Furnace  | Using Standard RTD, S-Type Sensor with Digital Thermometer by Single Position Calibration (At measuring location in DUC)                         | 200 °C to 600 °C  | 0.48°C   |
| 141  | THERMAL-TEMPERATURE | Temperature Indicator with sensor of Freezer, Chamber, Bath, Oven, & Furnace  | Using Standard RTD, S-Type Sensor with Digital Thermometer by Single Position Calibration (At measuring location in DUC)                         | -50 °C to 200 °C  | 0.2°C  |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

33 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|--|---|--|
| 142  | THERMAL-TEMPERATURE | Temperature Indicator with sensor of Freezer, Chamber, Bath, Oven, & Furnace  | Using S-Type Sensor with Digital Thermometer by Single Position Calibration (At measuring location in DUC) | 600 °C to 1000 °C   | 1.6°C  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

34 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No          | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                     | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---------------|---|---|--|---|--|
| Site Facility |   |   |  |   |  |
| 1             | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM ,By Direct Method                        | 1 mA to 10 mA   | 0.16%  |
| 2             | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM ,By Direct Method                        | 10 µA to 1 mA   | 0.85 % to 0.16 %                                 |
| 3             | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM ,By Direct Method                        | 10 mA to 10 A   | 0.16 % to 0.25 %                                 |
| 4             | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current (1 Phase) (At 50 Hz)   | Using HIOKI Digital Power Meter/Harmonic Analyzer,By Direct Method | 10 mA to 65 A   | 0.06%  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

35 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---|---|---|---|--|
| 5    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current (3 Phase) (At 50Hz)  | Using HIOKI Digital Power Meter/Harmonic Analyzer ,By Direct Method | 10 mA to 50 A   | 0.13 % to 0.034 %                                |
| 6    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)  | AC Current (1 Phase & 3 Phase) @ 50 Hz  | Using CALMET C300 Power/Energy Calibrator by Direct Method          | 20 A to 120 A   | 0.3%   |
| 7    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)  | AC Current (1 Phase) @ 50 Hz  | Using Fluke 9100 with Current coil ,By Direct Method                | 20 A to 1000 A  | 0.76%  |
| 8    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)  | AC Current (1 Phase) @ 50Hz to 5kHz   | Using Fluke 9100 MFC ,By Direct Method                              | 10 $\mu$ A to 300 $\mu$ A   | 4.1 % to 0.25 %                                  |
| 9    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)  | AC Current (1 Phase) @ 50Hz to 5kHz   | Using Fluke 9100 MFC ,By Direct Method                              | 300 mA to 20 A  | 0.25 % to 0.3 %                                  |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

36 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                                       | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--|---|---|---|--|
| 10   | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source) | AC Current(1 Phase) @ 50Hz to 5kHz  | Using Fluke 9100 MFC ,By Direct Method                              | 300 $\mu$ A to 300 mA   | 0.25%  |
| 11   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)               | DC CURRENT  | Using Fluke 8846A, 6½ DMM ,By Direct Method                         | 0.10 mA to 10 mA  | 0.085 % to 0.065 %                               |
| 12   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)               | DC CURRENT  | Using Fluke 8846A, 6½ DMM ,By Direct Method                         | 1 $\mu$ A to 100 $\mu$ A  | 3.06 % to 0.085 %                                |
| 13   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)               | DC CURRENT  | Using HIOKI Digital Power Meter/Harmonic Analyzer ,By Direct Method | 10 A to 65 A  | 0.13 % to 0.12 %                                 |
| 14   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)               | DC CURRENT  | Using Fluke 8846A, 6½ DMM ,By Direct Method                         | 10 mA to 10 A   | 0.065 % to 0.8 %                                 |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

37 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                              | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---|---|---|---|--|
| 15   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC CURRENT  | Using Fluke 9100 MFC, By Direct Method                                      | 10 $\mu$ A to 300 mA  | 0.59 % to 0.035 %                                |
| 16   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC CURRENT  | Fluke 9100 with current coils ,By Direct Method                             | 20 A to 1000 A  | 0.8 % to 0.8 %                                   |
| 17   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC CURRENT  | Using Fluke 9100 MFC, By Direct Method                                      | 300 mA to 20 A  | 0.035%   |
| 18   | ELECTRO-TECHNICAL-OTHERS (Measure)        | AC High Voltage (At 50Hz)   | Using H.V. Probe & DMM Fluke ,By Direct Method                              | 1 kV to 28 kV   | 2.33%  |
| 19   | ELECTRO-TECHNICAL-OTHERS (Measure)        | AC Power (1 Phase) (At 50Hz) At UPF (10V to 600 V 0.1A to 65 A)   | Using Yokogawa/HIOKI Digital power Meter/Harmonic Analyzer,By Direct Method | 0.1 W to 39 kW  | 0.9%   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

38 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|---|---|--|
| 20   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Power (3 Phase) (At 50Hz ,At UPF) (10 V to 600 V, 0.01A to 50 A)   | Using HIOKI Digital Power Meter/Harmonic Analyzer,By Direct Method  | 0.1 W to 30 kW  | 0.13 % to 0.034 %                                |
| 21   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Voltage (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM , By Direct Method                        | 1 mV to 100 mV  | 4.7 % to 0.12 %                                  |
| 22   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Voltage (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM , By Direct Method                        | 1 V to 1000 V   | 0.1%   |
| 23   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Voltage (1 Phase) (50 Hz to 5 kHz)   | Using Fluke 8846A, 6½ DMM , By Direct Method                        | 100 mV to 1 V   | 0.12 % to 0.1 %                                  |
| 24   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Voltage (1 Phase) (At 100 kHz)   | Using Fluke 8846A, 6½ DMM , By Direct Method                        | 10 mV to 100 V  | 6 % to 0.8 %                                     |
| 25   | ELECTRO-TECHNICAL-OTHERS (Measure) | AC Voltage (3 Phase) (At 50Hz)  | Using HIOKI Digital Power Meter/Harmonic Analyzer ,By Direct Method | 0.1 V to 1000 V   | 0.7 % to 0.06 %                                  |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

39 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                     | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|--|---|--|
| 26   | ELECTRO-TECHNICAL-OTHERS (Measure) | Capacitance (At 1 kHz)  | Using LCR Meter ,By Direct Method                                  | 1 $\mu$ F to 1011 $\mu$ F   | 0.91 % to 0.31 %                                 |
| 27   | ELECTRO-TECHNICAL-OTHERS (Measure) | Capacitance (At 1 kHz)  | Using LCR Meter ,By Direct Method                                  | 1 nF to 1000 nF   | 0.91 % to 0.06 %                                 |
| 28   | ELECTRO-TECHNICAL-OTHERS (Measure) | Capacitance (At 1 kHz)  | Using LCR Meter ,By Direct Method                                  | 1 pF to 1000 pF   | 0.96 % to 0.06 %                                 |
| 29   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC High Voltage   | Using H.V. Probe & DMM Fluke ,By Direct Method                     | 1 kV to 40 kV   | 2.38%  |
| 30   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC Power (10 V to 600 V, 0.01A to 65 A)   | Using HIOKI Digital Power Meter/Harmonic Analyzer,By Direct Method | 0.1 W to 39 kW  | 0.2%   |
| 31   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC Voltage  | Using Fluke 8846A, 6½ DMM ,By Direct Method                        | 0.1 mV to 1 mV  | 4.0 % to 0.4 %                                   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

40 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|--|---|--|
| 32   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC Voltage  | Using Fluke 8846A, 6½ DMM ,By Direct Method                                  | 1 mV to 100 mV  | 0.4 % to 0.0085 %                                |
| 33   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC Voltage  | Using Fluke 8846A, 6½ DMM , By Direct Method                                 | 1 V to 1000 V   | 0.0045 % to 0.006 %                              |
| 34   | ELECTRO-TECHNICAL-OTHERS (Measure) | DC Voltage  | Using Fluke 8846A, 6½ DMM , By Direct Method                                 | 100 mV to 1 mV  | 0.0085 % to 0.0045 %                             |
| 35   | ELECTRO-TECHNICAL-OTHERS (Measure) | Harmonics (1 Phase,3 Phase) (200V, 10A, 50Hz)   | Using Yokogawa/HIOKI Digital power Meter/Harmonic Analyzer ,By Direct Method | 3 <sup>rd</sup> to 49 <sup>th</sup> Order   | 0.51%  |
| 36   | ELECTRO-TECHNICAL-OTHERS (Measure) | Harmonics (1 Phase,3 Phase) (200V-240V, 0.5A to 10A, 50Hz)  | Using Yokogawa/HIOKI Digital power Meter/Harmonic Analyzer ,By Direct Method | 3 <sup>rd</sup> to 39 <sup>th</sup> Order   | 0.51%  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

41 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|--|---|--|
| 37   | ELECTRO-TECHNICAL-OTHERS (Measure) | Impulse (Duration - Half pulse width)   | Using Oscilloscope with HV Probe by Direct/Comparison method as per IEC 60060-2, IEC 61180, IS 2071 Part 1 | 10 $\mu$ S to 1000 $\mu$ S  | 1.6 % to 2.5 %                                   |
| 38   | ELECTRO-TECHNICAL-OTHERS (Measure) | Impulse (Front Time)  | Using Oscilloscope with HV Probe by Direct/Comparison method as per IEC 60060-2, IEC 61180, IS 2071 Part 1 | 0.8 $\mu$ S to 1000 $\mu$ S   | 1.6 % to 2.5 %                                   |
| 39   | ELECTRO-TECHNICAL-OTHERS (Measure) | Impulse (Voltage Magnitude)   | Using Oscilloscope with HV Probe by Direct/Comparison method as per IEC 60060-2, IEC 61180, IS 2071 Part 1 | 1 kVp to 17 kVp   | 3.69%  |
| 40   | ELECTRO-TECHNICAL-OTHERS (Measure) | Inductance (At 1kHz)  | Using LCR Meter ,By Direct Method  | 1 H to 10 H   | 0.061 % to 0.06 %                                |
| 41   | ELECTRO-TECHNICAL-OTHERS (Measure) | Inductance (At 1kHz)  | Using LCR Meter ,By Direct Method  | 1 mH to 1003 mH   | 0.06 % to 0.28 %                                 |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

42 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|--|---|--|
| 42   | ELECTRO-TECHNICAL-OTHERS (Measure) | Inductance (At 1kHz)  | Using LCR Meter ,By Direct Method  | 99 $\mu$ H to 999.9 $\mu$ H   | 0.28%  |
| 43   | ELECTRO-TECHNICAL-OTHERS (Measure) | Power Factor  | Using Yokogawa/HIOKI Digital power Meter/Harmonic Analyzer, By Direct Method | 0.1 1PF lag - UPF to 0.1 PF lead  | 0.007PF  |
| 44   | ELECTRO-TECHNICAL-OTHERS (Measure) | Resistance  | Using LCR Meter (At 1kHz), By Direct Method                                  | 0.01 Ohm to 100 Ohm   | 0.6 % to 0.07 %                                  |
| 45   | ELECTRO-TECHNICAL-OTHERS (Measure) | Resistance  | Using Fluke 8846A, 6½ DMM,By Direct Method                                   | 0.1 Ohm to 100 kOhm   | 3.6 % to 0.013 %                                 |
| 46   | ELECTRO-TECHNICAL-OTHERS (Measure) | Resistance  | Using Fluke 8846A, 6½ DMM,By Direct Method                                   | 10 MOhm to 1 GOhm   | 0.05 % to 2.5 %                                  |
| 47   | ELECTRO-TECHNICAL-OTHERS (Measure) | Resistance  | Using Fluke 8846A, 6½ DMM, By Direct Method                                  | 100 kOhm to 10 MOhm   | 0.013 % to 0.05 %                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

43 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------|---|--|---|--|
| 48   | ELECTRO-TECHNICAL-OTHERS (Measure) | Surge   | Using Oscilloscope with HV Probe by Direct/Comparison method as per IEC 60060-2, IEC 61180, IS 2071 Part 1 | . up to 10 kV   | 1.6 % to 3.69 %                                  |
| 49   | ELECTRO-TECHNICAL-OTHERS (Source)  | AC Energy (1Phase & 3Phase) (Active/Reactive) UPF to 0.25PF Lead/Lag @ 50Hz (63.5 V to 240 V, 0.05 A to 120 A)              | Using CALMET C300 Power/Energy calibrator ,By comparison method  | 0.79 Wh to 28.8 kWh   | 0.68 % to 0.06 %                                 |
| 50   | ELECTRO-TECHNICAL-OTHERS (Source)  | AC Power (1 Phase) (P.F. 0.8 Lag to 0.8 Lead) (At 50Hz)   | Using Fluke 9100 MFC,By Direct Method  | 2.4 W to 8.96 kW  | 0.08 % to 0.65 %                                 |
| 51   | ELECTRO-TECHNICAL-OTHERS (Source)  | AC Power (1 Phase) P.F. 0.5 Lag to 0.5 Lead (At 50Hz )  | Using Fluke 9100 MFC ,By Direct Method   | 1.5 W to 5.6 kW   | 0.08 % to 0.65 %                                 |
| 52   | ELECTRO-TECHNICAL-OTHERS (Source)  | AC Power (1 Phase) UPF @ 50Hz (30 V to 560 V, 0.1A to 20 A)   | Using Fluke 9100 MFC ,By Direct Method   | 3 W to 11.2 kW  | 0.06 % to 0.6 %                                  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

44 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|---|--|---|--|
| 53   | ELECTRO-TECHNICAL-OTHERS (Source) | AC Power (3 Phase) (UPF) @ 50Hz (10 V to 500 V, 0.1A to 100 A)  | Using CALMET C300 Power/Energy calibrator,By Direct Method   | 1 W to 50.0 kW  | 1.5%   |
| 54   | ELECTRO-TECHNICAL-OTHERS (Source) | AC Voltage (1 Phase) (At 10Hz to 100 KHz )  | Using Fluke 9100 MFC , By Direct Method                      | 10 mV to 32 mV  | 4.56 % to 0.55 %                                 |
| 55   | ELECTRO-TECHNICAL-OTHERS (Source) | AC Voltage (1 Phase) (10Hz to 100 kHz)  | Using Fluke 9100 MFC , By Direct Method                      | 32 mV to 320 mV   | 0.55 % to 0.25 %                                 |
| 56   | ELECTRO-TECHNICAL-OTHERS (Source) | AC Voltage (1 Phase) (10Hz to 100kHz)   | Using Fluke 9100 MFC , By Direct Method                      | 320 mV to 1000 V  | 0.25 % to 0.2 %                                  |
| 57   | ELECTRO-TECHNICAL-OTHERS (Source) | AC Voltage (3 Phase) (At 50 Hz)   | Using CALMET C300 Power/ Energy Calibrator ,By Direct Method | 0.5 V to 560 V  | 0.043 % to 0.049 %                               |
| 58   | ELECTRO-TECHNICAL-OTHERS (Source) | Capacitance (At 1 kHz)  | Using Fluke 9100 MFC , By Direct Method                      | 1 nF to 100 nF  | 2.04 % to 0.55 %                                 |
| 59   | ELECTRO-TECHNICAL-OTHERS (Source) | Capacitance (At 1 kHz)  | Using Fluke 9100 MFC , By Direct Method                      | 100 nF to 1 mF  | 0.55 % to 1.85 %                                 |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

45 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|---|--|---|--|
| 60   | ELECTRO-TECHNICAL-OTHERS (Source) | DC Power (1 Phase) (10 V to 600 V, 0.1A to 20 A)  | Using Fluke 9100 MFC,By Direct Method                            | 1 W to 12 kW  | 6.3 % to 0.2 %                                   |
| 61   | ELECTRO-TECHNICAL-OTHERS (Source) | DC Voltage  | Using Fluke 9100 MFC , By Direct Method                          | 1 mV to 300 mV  | 0.7 % to 0.055 %                                 |
| 62   | ELECTRO-TECHNICAL-OTHERS (Source) | DC Voltage  | Using Fluke 9100 MFC , By Direct Method                          | 100 V to 1000 V   | 0.015 % to 0.01 %                                |
| 63   | ELECTRO-TECHNICAL-OTHERS (Source) | DC Voltage  | Using Fluke 9100 MFC , By Direct Method                          | 300 mV to 100 V   | 0.055 % to 0.015 %                               |
| 64   | ELECTRO-TECHNICAL-OTHERS (Source) | High Resistance# (I.R Option) (at 250V to 1kV)  | Using Fluke 9100 MFC , By Direct Method                          | 1 MOhm to 1.8 GOhm  | 0.2 % to 1.1 %                                   |
| 65   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 0.1 mOhm to 0.1 mOhm  | 0.96%  |
| 66   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 1 mOhm to 1 mOhm  | 0.79%  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

46 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|---|--|---|--|
| 67   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 1 Ohm to 1 Ohm  | 0.31%  |
| 68   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 10 mOhm to 10 mOhm  | 0.31%  |
| 69   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 100 mOhm to 100 mOhm  | 0.31%  |
| 70   | ELECTRO-TECHNICAL-OTHERS (Source) | Low Resistance (4W)   | Using OSAW Standard Resistance (Direct Values) ,By Direct Method | 2 Ohm to 2 Ohm  | 0.31%  |
| 71   | ELECTRO-TECHNICAL-OTHERS (Source) | Power Factor @ 50Hz   | Using Fluke 9100 MFC by Direct Method                            | 0.1 PF lag -UPF to 0.1 PF lead  | 0.002 % to 0.001 %                               |
| 72   | ELECTRO-TECHNICAL-OTHERS (Source) | Resistance (2W)   | Using Fluke 9100 MFC ,By Direct Method                           | 1 MOhm to 10 MOhm   | 0.076 % to 0.20 %                                |
| 73   | ELECTRO-TECHNICAL-OTHERS (Source) | Resistance (2W)   | Using Fluke 9100 MFC ,By Direct Method                           | 10 Ohm to 100 kOhm  | 0.20 % to 0.035 %                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

47 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--|---|---|---|--|
| 74   | ELECTRO-TECHNICAL-OTHERS (Source)                  | Resistance (2W)   | Using Fluke 9100 MFC ,By Direct Method              | 10 MOhm to 300 MOhm   | 0.02 % to 0.59 %                                 |
| 75   | ELECTRO-TECHNICAL-OTHERS (Source)                  | Resistance (2W)   | Using Fluke 9100 MFC ,By Direct Method              | 100 kOhm to 1 MOhm  | 0.032 % to 0.076 %                               |
| 76   | ELECTRO-TECHNICAL-OTHERS (Source)                  | RESISTANCE (2W)   | Using Fluke 9100 MFC ,By Direct Method              | 3 Ohm to 10 Ohm   | 0.46 % to 0.20 %                                 |
| 77   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | Temperature Simulation (By simulation method) (J Type Thermocouple)   | Using Cropico digital thermometer, By Direct Method | -200 °C to 1200 °C  | 0.55°C   |
| 78   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | Temperature Simulation (By simulation method) (K Type Thermocouple))  | Using Cropico digital thermometer, By Direct Method | -200 °C to 1300 °C  | 0.52°C   |
| 79   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | Temperature Simulation (By simulation method) (RTD Type)  | Using Cropico digital thermometer, By Direct Method | -200 °C to 600 °C   | 0.2°C  |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

48 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---|---|--|---|--|
| 80   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | Harmonics (1Phase & 3Phase) (10V to 300V, 0.5A to 100A)   | Using CALMET C300 Power/Energy Calibrator , By Direct Method | 2 <sup>nd</sup> to 31 <sup>st</sup>   | 0.6%   |
| 81   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | Temperature Indicator, Controller, Recorders (By simulation method) (J Type Thermocouple)                                   | Using Fluke 9100 MFC, By Direct Method                       | -200 °C to 1200 °C  | 0.3°C  |
| 82   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | Temperature Indicator, Controller, Recorders (By simulation method) (RTD TYPE)  | Using Fluke 9100 MFC, By Direct Method                       | -200 °C to 800 °C   | 0.2°C  |
| 83   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | Temperature Indicator, Controller, Recorders (By simulation method) (K Type Thermocouple)                                   | Using Fluke 9100 MFC, By Direct Method                       | -200 °C to 1300 °C  | 0.3°C  |
| 84   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | FREQUENCY (1 Phase)   | Using Fluke 8846A, 6½ DMM ,By Direct Method                  | 10 Hz to 1 MHz  | 0.17 % to 0.011 %                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

49 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                           | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--|---|---|---|--|
| 85   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure) | FREQUENCY (3 Phase) (At 240V)   | Using HIOKI Digital Power Meter/Harmonic Analyzer,By Direct Method  | 50 Hz to 50 kHz   | 0.02%  |
| 86   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure) | TIME INTERVAL   | Using Time Interval Meter & Digital Stop Watch,By Comparison Method | 0.1 s to 999.99 s   | 0.002 Sec to 1.13 Sec                            |
| 87   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure) | TIME INTERVAL   | Using Time Interval Meter & Digital Stop Watch,By Comparison Method | 1000 s to 24 hrs  | 0.002 sec to 1.61 sec                            |
| 88   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)  | FREQUENCY   | Using Fluke 9100 MFC , By Direct Method                             | 1 Hz to 100 Hz  | 0.6 % to 0.006 %                                 |
| 89   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)  | FREQUENCY   | Using Fluke 9100 MFC , By Direct Method                             | 100 Hz to 1 MHz   | 0.006 % to 0.003 %                               |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

50 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---|---|---|---|--|
| 90   | MECHANICAL-PRESSURE INDICATING DEVICES                    | (Pneumatic-Gauge Pressure) Digital & Dial Pressure Gauge  | Using Digital Pressure Gauge & Comparator by Comparison Method as per DKD- R - 6 -1   | 0 to 10 bar   | 0.20% of rdg                                     |
| 91   | MECHANICAL-PRESSURE INDICATING DEVICES                    | (Pneumatic-Gauge Pressure) Digital & Dial Pressure Gauge  | Using Digital Vacuum Gauge & Vacuum Comparator by Comparison Method as per DKD- R - 6 -1  | -90 bar to 0 bar  | 0.01% of rdg                                     |
| 92   | MECHANICAL-PRESSURE INDICATING DEVICES                    | Hydraulic Pressure # ( Analog / Digital Gauges)   | Using Digital Pressure Gauge with Hydraulic Comparator by Comparison Method as per DKD- R - 6 -1                                | 0 to 700 bar  | 0.1% of rdg                                      |
| 93   | MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE | Force Measuring System of UTM Compression   | Using Load cell with indicator of Class 0.5 and Class 1 accuracy by UTM of accuracy Class I and coarser based on ISO 1828,Part1 | 1 kN to 10 kN   | 0.24%  |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

51 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---|---|--|---|--|
| 94   | MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE | Force Measuring System of UTM Tension   | Using Load cell with indicator of Class 0.5 and Class 1 accuracy by UTM of accuracy Class I and coarser based on ISO 1828,Part1  | 1 kN to 10 kN   | 0.59%  |
| 95   | MECHANICAL-WEIGHING SCALE AND BALANCE                     | Mass-Electronic weighing balances with readability d=10g  | Using M1 class weights by Calibration of electronic weighing balance and comparator of Class III and coarser as per OIML R-76- 1 | Maximum capacity up to 10 to  | 12.9g  |
| 96   | MECHANICAL-WEIGHING SCALE AND BALANCE                     | Mass-Electronic weighing balances with readability d=10g  | Using M1 class weights by Calibration of electronic weighing balance and comparator of Class III and coarser as per OIML R-76- 1 | Maximum capacity up to 20   | 1.2g   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

52 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                    | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------|---|--|---|--|
| 97   | MECHANICAL-WEIGHING SCALE AND BALANCE | Mass-Electronic weighing balances with readability d=10g  | Using M1 class weights by Calibration of electronic weighing balance and comparator of Class III and coarser as per OIML R-76- 1     | Maximum capacity up to 60   | 3g   |
| 98   | MECHANICAL-WEIGHING SCALE AND BALANCE | Mass-Electronic weighing balances with readability d=1mg  | Using E2 class weights by Calibration of electronic weighing balance and comparator of Class I and coarser as per OIML R-76- 1       | Maximum capacity up to 20   | 1mg  |
| 99   | MECHANICAL-WEIGHTS                    | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 1 g   | 0.11mg   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

53 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 100  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 1 mg  | 0.11mg   |
| 101  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 10 g  | 0.11mg   |
| 102  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 10 mg   | 0.11mg   |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

54 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 103  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 100 g   | 0.12mg   |
| 104  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 100 mg  | 0.11mg   |
| 105  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 2 g   | 0.11mg   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

55 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 106  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 2 mg  | 0.11mg   |
| 107  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 20 g  | 0.12mg   |
| 108  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 20 mg   | 0.11mg   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

56 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 109  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 200 g   | 0.16mg   |
| 110  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 200 mg  | 0.11mg   |
| 111  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 5 g   | 0.11mg   |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

57 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|--|---|--|
| 112  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 5 mg  | 0.11mg   |
| 113  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 50 g  | 0.12mg   |
| 114  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111 | 50 mg   | 0.11mg   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

58 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|---|---|---|--|
| 115  | MECHANICAL-WEIGHTS | Mass-Weights  | Using E2 class weights and balance of readability 0.1mg by Calibration of weights of Class M1 accuracy and coarser as per OIML R-111  | 500 mg  | 0.11mg   |
| 116  | OPTICAL-OPTICAL    | Chromaticity Coordinates of LED Lamp  | Using C-Type mirror Goniometer with Spectroradiometer and integrating sphere with standard spectroradiometer/ Spectrophotometer and Standard light source by Obsolete & Substitution method | x =0.005  | 1.0%   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

59 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|---|---|--|
| 117  | OPTICAL-<br>OPTICAL | Chromaticity Coordinates of LED Lamp  | Using C-Type mirror Goniometer with Spectroradiometer and integrating sphere with standard spectroradiometer/ Spectrophotometer and Standard light source by Obsolete & Substitution method | y =0.005  | 1.0%   |
| 118  | OPTICAL-<br>OPTICAL | Color Coordinates   | Using THL lamps at 50W & integrating sphere with standard spectrophotometer by ab-initio Method & Substitution method   | x =0.005  | 1.0%   |
| 119  | OPTICAL-<br>OPTICAL | Color Coordinates   | Using THL lamps at 50W & integrating sphere with standard spectrophotometer by ab-initio Method & Substitution method   | y =0.005  | 1.0%   |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

60 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|---|---|--|
| 120  | OPTICAL-<br>OPTICAL | Compact Fluorescent Lamps (CFL) (5 W to 26 W)   | Using standard lamps & integrating sphere with standard spectrophotometer by ab-initio Method & Substitution method   | 200 lm to 1600 lm   | 3.0%   |
| 121  | OPTICAL-<br>OPTICAL | Correlated Color Temperature  | Using THL lamps at 50W & integrating sphere with standard spectrophotometer by ab-initio Method & Substitution method   | 1500 K to 25000 K   | 1.5%   |
| 122  | OPTICAL-<br>OPTICAL | Correlated Color Temperature of LED Lamp  | Using C-Type mirror Goniometer with Spectroradiometer and integrating sphere with standard spectroradiometer/ Spectrophotometer and Standard light source by Obsolete & Substitution method | 1500 K to 25000 K   | 1.2%   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

61 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group                   | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------------------------|---|--|---|--|
| 123  | OPTICAL-<br>OPTICAL                  | Illuminance   | Using Illuminance meter with Intensity Standard Lamp WI 41G Lamp by Direct Method  | 1 lux to 25000 lux  | 2.1%   |
| 124  | THERMAL-<br>SPECIFIC HEAT & HUMIDITY | Calibration of Furnace, Freezer, Oven ,Chamber  | Using 16-Channel Temperature Scanner with minimum 9 RTD Sensors by Multi Position Calibration  | 140 °C to 250 °C  | 1.7°C  |
| 125  | THERMAL-<br>SPECIFIC HEAT & HUMIDITY | Calibration of Furnace, Freezer, Oven ,Chamber  | Using 16-Channel Temperature Scanner with minimum 9 RTD Sensors by Multi Position Calibration  | -40 °C to 140 °C  | 1.7°C  |
| 126  | THERMAL-<br>TEMPERATURE              | Calibration of Contact type temp. sensor (RTDs, Thermocouples, Temp. indicators with Sensors)                               | Using Standard RTD, S- Type T/C Sensor with Standard Digital Thermometer and Low Temperature Bath & Dry Block Calibrators , By Comparison Method | 100 °C to 200 °C  | 0.14°C   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

62 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|--|---|--|
| 127  | THERMAL-TEMPERATURE | Calibration of Contact type temp. sensor (RTDs, Thermocouples, Temp. indicators with Sensors)                               | Using Standard RTD, S- Type T/C Sensor with Standard Digital Thermometer and Low Temperature Bath & Dry Block Calibrators , By Comparison Method | 200 °C to 600 °C  | 0.32°C   |
| 128  | THERMAL-TEMPERATURE | Calibration of Contact type temp. sensor (RTDs, Thermocouples, Temp. indicators with Sensors)                               | Using Standard RTD, S- Type T/C Sensor with Standard Digital Thermometer and Low Temperature Bath & Dry Block Calibrators , By Comparison Method | -40 °C to 100 °C  | 0.11°C   |
| 129  | THERMAL-TEMPERATURE | Calibration of Contact type temp. sensor (Thermocouples, Temp. indicators with Sensors)                                     | Using Standard RTD, S- Type T/C Sensor with Standard Digital Thermometer and Low Temperature Bath & Dry Block Calibrators , By Comparison Method | 600 °C to 1000 °C   | 1.55°C   |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2165

**Page No**

63 of 64

**Validity**

03/12/2018 to 02/12/2020\*

**Last Amended on**

-

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---|--|---|--|
| 130  | THERMAL-TEMPERATURE | Glass Thermometer, Dial Gauge   | Using Standard RTD Sensor with Standard Digital Thermometer and Low Temperature Bath , By Comparison Method              | -40 °C to 200 °C  | 0.65°C   |
| 131  | THERMAL-TEMPERATURE | Temperature Indicator with sensor of Freezer, Chamber, Bath, Oven, & Furnace  | Using Standard RTD, S-Type Sensor with Digital Thermometer by Single Position Calibration (At measuring location in DUC) | 200 °C to 600 °C  | 0.48°C   |
| 132  | THERMAL-TEMPERATURE | Temperature Indicator with sensor of Freezer, Chamber, Bath, Oven, & Furnace  | Using Standard RTD, S-Type Sensor with Digital Thermometer by Single Position Calibration (At measuring location in DUC) | -50 °C to 200 °C  | 0.2°C  |
| 133  | THERMAL-TEMPERATURE | Temperature Indicator with sensor of Freezer, Chamber, Bath, Oven, & Furnace  | Using S-Type Sensor with Digital Thermometer by Single Position Calibration (At measuring location in DUC)               | 600 °C to 1000 °C   | 1.6°C  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

|                               |   |                        |          |
|-------------------------------|---|------------------------|----------|
| <b>Laboratory Name :</b>      | HI PHYSIX LABORATORY INDIA PRIVATE LIMITED, B-32/1/2, M.I.D.C. INDUSTRIAL AREA , RANJANGAON, PUNE, MAHARASHTRA, INDIA |                        |          |
| <b>Accreditation Standard</b> | ISO/IEC 17025:2017  |                        |          |
| <b>Certificate Number</b>     | CC-2165   | <b>Page No</b>         | 64 of 64 |
| <b>Validity</b>               | 03/12/2018 to 02/12/2020*   | <b>Last Amended on</b> | -        |

\* The validity is extended for one year up to 02.12.2021

\*Transition to 2017 version completed w.e.f 19.03.2021 valid until 02.12.2021

\* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of  $k = 2$ .

